



दूरभाष : 01437-220177
फैक्स नं. 91-01437-220163
भा.कृ.अ.प.—केन्द्रीय भेड़ एवं ऊन अनुसंधान संस्थान
अविकानगर, तह0 मालपुरा, जिला-टोंक (राजस्थान) - 304501
ICAR-Central Sheep & Wool Research Institute
Avikanagar, Teh.Malpura, Dist.Tonk (Rajasthan) - 304501



F.No.1(387)SP/PAC/2010/V.III/

Dated: 25.08.2018

Corrigendum


The Director, ICAR-CSWRI, Avikanagar invited online bids in two bid system from reputed manufacturer firms and their authorized dealers, Indian agent on behalf of foreign principals for supply of Scientific Equipment i.e. Fluorescence Microscope vide tender ID No. 2018_DARE_368388_1 dated 03.08.2018.

The technical specification of the equipment stands amended as below :-

S.No.	Tender Specification	Revised specification
2	Illumination: LED source not less than 4W, uniform light distribution, intensity controller, life span more than 40,000 hrs.	Illumination: LED source with power consumption not less than 3W, uniform light distribution, intensity controller, life span more than 40,000 hrs.
9.	Camera: Coloured scientific CCD (sensor chip size 2/3") or C-mos camera (6 x 4mm sensor size or higher) with peltier cooling to capture BF/PH/Fluorescence images, Resolution 5MP, 1024 x 1024 resolutions at speed/live display of 30 fps or better one. Bit depth and dynamic range should be higher for quality image acquisition. Should have USB 3.0 PC interface and include appropriate c-mount adopter. Suitable optical mount should be provided along with licensed image analysis software from same manufacturer for 2D/3D imaging, spatial measurements such as length, width, area, perimeter, counting, intensity profiling, live image capture, video recording, channel merging, annotations, image gallery and image comparison, crop and image arithmetic and other advanced features. Microscope, Objective, camera and software should be from same manufacturer for better integration and performance.	Camera: Coloured scientific CCD (sensor chip size 2/3") or C-mos camera (6 x 4mm sensor size or higher) to capture BF/PH/Fluorescence images, Resolution 5MP, 1024 x 1024 resolutions at speed/live display of 30 fps or better one. 12 Bit depth and dynamic range should be higher for quality image acquisition. Should have USB 3.0 PC interface and include appropriate c-mount adopter. Suitable optical mount should be provided along with licensed image analysis software from same manufacturer for 2D/3D imaging, spatial measurements such as length, width, area, perimeter, counting, intensity profiling, live image capture, video recording, channel merging, annotations, image gallery and image comparison, crop and image arithmetic and other advanced features. Microscope, Objective, camera and software should be from same manufacturer for better integration and performance.
14	Certification: USFDA enlistment, ISO, European C.	Certification: USFDA enlistment/ European C.

The last date of bid submission is extended till 04.09.2018 at 3.00 PM and the date of tender opening is 05.09.2018 at 3.00 PM. Other terms & conditions will be the same.

This is issued with the approval of competent authority.


(Harshit Agarwal)
Administrative Officer